

Setup Guide for EVAsoft/ProImage 3.4 and higher.

 Install the REALCLOUD HD1 camera button software utility. The installer is located on the REALCLOUD HD1 flash drive in the Button Software folder. Or REALCLOUD Website <u>http://www.realcloudimaging.com/support/</u> Scroll down the Support Page to the Downloads Section. Click the "Capture Button Software"

button. This will start downloading the Button-Tray-Utility.zip , once downloaded open the zip file and Install the REALCLOUD HD1 camera button tray utility software.

2) Select the Next button on the Welcome screen.





3) Select the install destination folder and click on the Install button.

🛃 Button Tray Utility 2.1 Setup	
Choose Install Location Choose the folder in which to install Button Tray Utility 2.1.	L
Setup will install Button Tray Utility 2.1 in the following folder. To install in a di click Browse and select another folder. Click Install to start the installation.	fferent folder,
Destination Folder C:\Program Files (x86)\Button Tray Utility B	<u>r</u> owse
Space required: 39.6 MB Space available: 753.7 GB	
Button Tray Utility 2.1.6919.19922 <u>Sack</u> Install	Cancel

4) Plug the REALCLOUD HD1 camera into the computer and click on the OK button to start the installation prosses.





5) Click on the Next button to continue.


Button Tray Utility 2.1 Setup	
Installation Complete Setup was completed successfully.	
Completed	
Show <u>d</u> etails	
Button Tray Utility 2,1,6919,19922	
	< <u>B</u> ack <u>Next</u> Cancel



6) Select a reboot option (if prompted) and click on the Finish button.

Button Tray Utility 2.1 Setup	
	Completing Button Tray Utility 2.1 Setup
	Your computer must be restarted in order to complete the installation of Button Tray Utility 2.1. Do you want to reboot now?
	Reboot now
凤	I want to manually reboot later
	< Back Finish Cancel

7) Right click on the Button Tray Utility icon in your system tray and select Preferences.





8) Click in the first text box for Short Press Keys and press the "spacebar" key one time on your keyboard, then click on the Done button.

Note: The second text box for Short Press Keys and the text box for Long Press Keys are left blank.

Preferences	×	
Short Press Keys:		
Status Camera Registered Camera Streaming Button Click		
Done		




9) Open the EVAsoft Configuration Tool application, select the Mode tab, place a checkmark next to the Enable Camera Devices option and click on the Save button.


🛃 Configuration Tool	_ _ ×
File Help	
Mode EVA EVA SELECT Flashscan Panoramic Twain	
Select the features to Enable/Disable:	
Debugger Mode	
Display in System Tray when in Network Host Mode	
☑ Enable Camera Devices	
Enable TWAIN Devices	
Legacy ProImage DDE Mode	
Hide Chart View Tab	
For user with the ImageV PACS Software ONLY:	
Protice Management Mode	
Ca	ncel Save



10) Open the EVAsoft imaging application, click on the Image Device file menu and select RCI 1080P HD.



11) With a patient selected, click on the Acquire tab and select the Start Series Exposure Mode button to open the video acquisition.





12) A short/quick camera button press, and release will save an image from the REALCLOUD HD1 camera.

NOTE: EVAsoft imaging application does not have an option/setting to open any camera property page, if you need to make camera adjustments (brightness, contrast, hue, mirror...) you will need to make the adjustment in another application (AmCap).